



# Chandigarh State AIDS Control Society

International Hostel (Near Congress Bhawan), Sector 15-A, Chandigarh - 160015

Phone Nos. 0172-2783300 Fax No. 0172-2700171 or E-mail us at [chandigarhsacs@gmail.com](mailto:chandigarhsacs@gmail.com)

S.No	Item	Product Description/Specification	QTY
1.	Safety Lancets	Sterile single use disposable auto retractable safety lancet having stainless steel needle (21 gauge & 2.0-2.2 mm depth)	2500
2.	Dry Blood Spot Card (Protein Saver Card)	<ul style="list-style-type: none"> <li>• No of circles: Minimum five circles</li> <li>• Circle size: t2-L4 mm internal Diameter</li> <li>• Dimensions: Height: 3" (76.2 mm), Length 4" (108 mm)</li> <li>• Basic weight (g/m<sup>2</sup>) : 170-188</li> <li>• Expiry: Min 3 years (36 months) at the time of delivery</li> <li>• pH range should be between 5.7 to 7.5</li> <li>• Certification of product: US FDA or CE</li> <li>• Packing: 10 DBS cards per pack</li> <li>• Other information:               <ol style="list-style-type: none"> <li>1. Labelling space should be available for (a) sample code and (b) date of collection                   <ul style="list-style-type: none"> <li>○ The card is imprinted with the universal biohazard symbol</li> </ul> </li> </ol> </li> </ul>	2500
3.	Drying racks for DBS Cards	Foldable vertical drying rack made of paper card, without Velcro and individually sealed in zip-lock bag, allowing air drying of Dried Blood specimen in horizontal position, with a capacity to keep 10 DBS cards at a time; the approximate size of 9 cm width, 20-22 cm height with an angular cut (25-40 degrees) for each drying slot of 5 cm length on both side with central part of 2.5 cm; drying rack card thickness of minimum 0.5 mm to ensure stability. Folding instructions should be printed on the back side of the rack.	30
4.	Silica Gel Packs	Silica Gel Pack (Silica Gel Sachet 1gm, complete with opaque non gas permeable bag in zip lock bag of approx. 150/180 mm size)	10000

Quoted rates must be inclusive of all taxes and delivery charges or any other charges applicable must be mentioned in the quotation.

You are requested to provide your quotations latest by 2/10/2023, 05:00 PM.

1 sample of each item must be provided by the L1 bidder before complete delivery.

We look forward to receiving your quotations and thank you for your interest in this Project.